

| | Analyte | Sample Volume - Minimum* | Sample Volume - Desired |
|------------|--|--------------------------|-------------------------|
| Organics | Volatiles (5035), Gasoline Range Organics- MI | 10 g | 10g |
| | Volatiles (5030), Gasoline Range Organics- OH | 10 g | 4 oz |
| | Semi Volatiles | 30 g | 4 oz |
| | PCBs | 15 g | 4 oz |
| | Organochlorine Pesticides | 15 g | 4 oz |
| | Organochlorine Pesticides/ PCBs | 15 g | 4 oz |
| | 2,4-D, 2,4,5-TP | 30 g | 4 oz |
| | Organochlorine Herbicides | 30 g | 4 oz |
| | Diesel Range Organics | 15 g | 4 oz |
| | Ethylene Glycol | 10 g | 4 oz |
| | Glycols (ethylene, propylene, triethylene) | 10 g | 4 oz |
| | Alcohols (n-butanol, ethanol, isobutanol, isopropanol, methanol) | 10 g | 4 oz |
| | Analyte | Sample Volume - Minimum* | Sample Volume - Desired |
| | Metals | 50 g | 4 oz |
| Inorganics | Mercury | 50 g | 4 oz |
| | Bromide | 5 g | 4 oz |
| | Chromium, Hexavalent | 15 g | 4 oz |
| | Chloride | 5 g | 4 oz |
| | Corrosivity | 5 g | 4 oz |
| | Cyanide, Total | 5 g | 4 oz |
| | Fluoride | 5 g | 4 oz |
| | Ignitability | 10 g | 4 oz |
| | Nitrogen, Ammonia | 5 g | 4 oz |
| | Nitrogen, Total Kjeldahl | 10 g | 4 oz |
| | Nitrogen, Nitrate | 5 g | 4 oz |
| | Nitrogen, Nitrite | 5 g | 4 oz |
| | Nitrate Plus Nitrite | 5 g | 4 oz |
| | pH | 5 g | 4 oz |
| | Phosphorus (water soluble) | 5 g | 4 oz |
| | Sulfate | 5 g | 4 oz |
| | Sulfide, Total | | 4 oz |
| Misc | Analyte | Sample Volume - Minimum | Sample Volume - Desired |
| | % Moisture | 30 g | 4 oz |

* Bare Minimum sample volumes do not account for the percent moisture sample volume

| | Analyte | Sample Volume - Minimum | Sample Volume - Desired |
|------------|--|-------------------------|-------------------------|
| Organics | Volatiles, Gasoline Range Organics | 40 mL | 120 mL |
| | Semi Volatiles | 650 mL | 1 Liter |
| | PCBs | 650 mL | 1 Liter |
| | Organochlorine Pesticides | 650 mL | 1 Liter |
| | Organochlorine Pesticides/ PCBs | 650 mL | 1 Liter |
| | 2,4-D, 2,4,5-TP | 650 mL | 1 Liter |
| | Organochlorine Herbicides | 650 mL | 1 Liter |
| | Diesel Range Organics | 650 mL | 1 Liter |
| | Ethylene Glycol | 2 mL | 40 mL |
| | Glycols (ethylene, propylene, triethylene) | 2 mL | 40 mL |
| | Alcohols (n-butanol, ethanol, isobutanol, isopropanol, methanol) | 2 mL | 40 mL |
| | Analyte | Sample Volume - Minimum | Sample Volume - Desired |
| | Mercury | 100 mL | 250 mL |
| | Metals Filtering (dissolved metals) | 100 mL | 250 mL |
| Inorganics | Total Metals | 100 mL | 250 mL |
| | Alkalinity | 50 mL | 250 mL |
| | Biochemical Oxygen Demand | 300 mL | 500 mL |
| | Bromide | 10 mL | 50 mL |
| | Carbon, Total Organic | 10 mL | 40 mL VOA |
| | Chemical Oxygen Demand | 5 mL | 10 mL |
| | Chromium, Hexavalent | 50 mL | 125 mL |
| | Chloride | 10 mL | 50 mL |
| | Chlorine Demand | 2 Liters | 2 Liters |
| | Chlorine, Total Residual | 25 mL | 50 mL |
| | Cyanide, Amenable | 50 mL | 250 mL |
| | Cyanide, Available | 20 mL | 40 mL |
| | Cyanide, Total | 50 mL | 250 mL |
| | Dissolved Oxygen | 50 mL | 125 mL |
| | Fats, Oils, & Grease | 100 mL | 500 mL |
| | Fluoride | 10 mL | 50 mL |
| | Hardness | 100 mL | 250 mL |
| | Ignitability | 50 mL | 125 mL |
| | Low Level Mercury | 40 mL | 120 mL |
| | Nitrogen, Ammonia | 100 mL | 250 mL |
| | Nitrogen, Total Kjeldahl | 50 mL | 250 mL |
| | Nitrogen, Nitrate | 10 mL | 50 mL |
| | Nitrogen, Nitrite | 10 mL | 50 mL |
| | Nitrate Plus Nitrite | 10 mL | 50 mL |
| | pH | 10 mL | 25 mL |
| | Phenolics, Total | 100 mL | 250 mL |
| | Phosphorus, Ortho (dissolved)* | 100 mL | 250 mL |
| | Phosphorus, Total | 100 mL | 250 mL |
| | Reactivity | 25 mL | 125 mL |
| | Solids, Total | 50 mL | 250 mL |
| | Solids, Total Dissolved | 50 mL | 250 mL |
| | Solids, Total Suspended | 50 mL | 250 mL |
| | Solids, Total Volatile | 50 mL | 250 mL |
| | Specific Conductance | 25 mL | 50 mL |
| | Sulfate | 10 mL | 50 mL |
| | Sulfide, Total | 50 mL | 250 mL |
| | Turbidity | 100 mL | 250 mL |